





Recycling Passport

Product name:	OPTIMUS 65 C 1T	
Identification code	9890 000 61571	
Order No.		
Device Type	B2B (Business to business)	
Producer /Manufacturer	Company name:	Philips Medical Systems DMC GmbH
	Address:	Roentgenstrasse 24
	Zip code:	22335 Hamburg
	Country:	Germany
	Web info:	http://staging.medical.philips.com/main/company/sustainability/recycling/
Total weight (in Kg) thereof		233

steel, iron	iron, low alloy (<5%)	105,429	KG
	iron, high alloy (>5%)	0,588	KG
steel, iron		106,017	KG
nonferrous metals and alloys	copper, -alloy	3,925	KG
	aluminum, -alloy	0,999	KG
	zinc, -alloy	0,438	KG
nonferrous metals and alloys		5,362	KG
glass / ceramics	glass	0,288	KG
glass / ceramics		0,288	KG
plastics / organic substances	oil	44,319	KG
	thermoplastic	5,161	KG
	thermoset	1,061	KG
	elastomer	0,362	KG
plastics / organic substances		50,903	KG
standard parts	other electronic powered devices	48,983	KG
	printed circuit boards	9,505	KG
	cables	7,596	KG
	mounting parts, attaching part	2,054	KG
	monitors	0,688	KG
	motors, pumps	0,52	KG
	fans	0,449	KG
	computers and accessories	0,407	KG
standard parts		70,202	KG
			232,772

Recycling Passport

Recycle info		
General Hazards (Desinfection)	For dismantling activities Treatment Facilities must consider the national requirements. For personnel which can come into contact with contaminated material, preventive measures pursuant to national requirements must be taken into account	
Hazardous Substances  To be removed	Substances:	Location
	Contact springs: Copperberyllium CuBe	Figure 2
Batteries  To be removed	Type:	Location
	Lithium chrome cell, 3V	Figure 1 (1)
	Lithium chrome cell, 3V	Figure 1 (2)
Special attention 	Item	Location
	Pb is present in the soldering process of PCBs	Figure 1, 2
Fluids / Gases 	Item	Location
	Transformer Oil, Shell Diala	Figure 1 (3)
	HV-transformer oil contains no PCBs	

Recycling Passport





Figure 1



Contact springs: Copperberyllium

Figure 2