





Recycling Passport

Product name:	ROT 5001 / ROT 5002 / ROT 5003		
	MRC xxx CT	MRC xxx	
Identification code	9890-000-63571	9890-000-85891	9890-000-85921
	9890-000-03782		
	9890-000-63481		
Order No.			
Device Type	B2B (Business to business)		
Producer /Manufacturer	Company name:	Philips Medical Systems DMC GmbH	
	Address:	Roentgenstrasse 24	
	Zip code:	22335 Hamburg	
	Country:	Germany	
	Web info:	http://staging.medical.philips.com/main/company/sustainability/recycling/	
Total weight (in Kg) thereof			58,0

steel, iron	iron, low alloy (<5%)	11,4KG
	iron, high alloy (>5%)	6,9KG
steel, iron		18,3KG
nonferrous metals and alloys	molybdenum and -alloys	7,5KG
	aluminium and -alloys	4,6KG
	copper and -alloys	2,8KG
nonferrous metals and alloys		14,9KG
glass / ceramics	ceramics	1,6KG
glass / ceramics		1,6KG
plastics / organic substances	oil	7,8KG
	thermoset	4,6KG
	elastomer	0,7KG
plastics / organic substances		13,1KG
relevant materials	lead and -alloys	8,3KG
	cobalt and -alloys	0,9KG
	tungsten and -alloys	0,3KG
	beryllium and -alloys	0,01KG
relevant materials		9,51KG

Printed copies are uncontrolled documents


Recycling Passport

Recycle info		
General Hazards (Desinfection)	For dismantling activities Treatment Facilities must consider the national requirements. For personnel which can come into contact with contaminated material, preventive measures pursuant to national requirements must be taken into account	
Hazardous Substances  To be removed	Substances:	Location
	Lead as x-ray shielding inside housing: 8,3 kg	Figure 1
	Beryllium as x-ray window in vacuum envelope: 10g	Figure 1
Batteries  To be removed	Type:	Location
Special attention 	Item	Location
Fluids / Gases 	Item	Location
	Oil: 7,8 kg	
	HV-transformer oil contains no PCBs	

Printed copies are uncontrolled documents

Recycling Passport

Locations as mentioned in the Passport (Pictures information)



Printed copies are
uncontrolled documents

Figure 1