## **Recycling Passport**

Product name:	OPTIMUS 65/80 R 1T CAN 480V		
<b>Identification code</b>	9890-000-61735		
Order No.			
Device Type	B2B (Business to business)		
Producer /Manufacturer	Company name:	Philips Medical Systems DMC GmbH	
Avianuracturer	Address:	Roentgenstrasse 24	
	Zip code:	22335 Hamburg	
	Country:	Germany	
	Web info:	http://www.medical.philips.com/main/company/sustainability/recycling/xray_generators/	
Total weight (in K	g)	264	
thereof			

ataal iran	iron low alloy ( <e0 )<="" th=""><th>00.460</th><th>VC.</th></e0>	00.460	VC.		
steel, iron	iron, low alloy (<5%)	99,469	KG		
	iron, high alloy (>5%)	1,706	KG		
steel, iron		101,175	KG		
nonferrous metals and alloys	copper, -alloy	3,959	KG		
	aluminum, -alloy	0,786	KG		
	zinc, -alloy	0,438	KG		
nonferrous metals and alloys	5,183	KG			
glass / ceramics	glass	0,288	KG		
glass / ceramics 0,288					
plastics / organic substances	oil	50,069	KG		
	thermoplastic	5,304	KG		
	thermoset	3,137	KG		
	elastomer	0,221	KG		
plastics / organic substances		58,731	KG		
	other electronic powered				
standard parts	devices	81,483	KG		
	printed circuit boards	9,759	KG		
	cables	5,486	KG		
	mounting parts, attaching part	2,187	KG		
	fans	0,064	KG		
standard parts		98,979	KG		
Not relevant for substances list	No relevant for subst. list				
Not relevant for substances list					
		264,356	KG		

SAP-Doc. 4512-105-06961   Doctype: 129-M03   S	SAP-Vers.: 00	Date :07-03-2007	Page1/3
--	---------------	------------------	---------

## **Recycling Passport**

Recycle info					
General Hazards (Desinfection)	eneral Hazards For dismantling activities Treatment Facilities must				
Hazardous Substances  To be removed	Substances:	Location			
Batteries  To be removed	Lithium chrome cell, 3V Lithium chrome cell, 3V	Location Figure 1 (1) Figure 1 (2)			
Special attention	Item Pb is present in the soldering process of PCBs	Location Figure 1, 2			
Fluids / Gases	Transformer Oil, Shell Diala  HV-transformer oil contains no PCBs	Location Figure 1 (3)			

## **Recycling Passport**




Figure 1



Figure 2