

Philips Medical Systems
DICOM Conformance Statement
CT Tomoscan M/EG/EG Compact
Release 3.1

Document Number 9896 020 05211
5 November 1998

© Copyright Philips Medical Systems Nederland B.V. 1998

All rights reserved

Issued by:

Philips Medical Systems Nederland B.V.

Integrated Clinical Solutions, Marketing & Communications

Building QP-0233

P.O. Box 10.000

5680 DA Best

The Netherlands

Tel.: +31 40 2763827

Fax.: +31 40 2763810

email: dicom@best.ms.philips.com

Internet (with the latest versions of Conformance Statements and other DICOM information):

<http://www.philips.com/ms/solution/connect>

ftp://ftp.philips.com/pub/ms/dicom/Conformance_Stmnts

Table of Contents

1	Introduction	5
1.1	Scope and field of application	5
1.2	Intended audience	5
1.3	Contents and structure	5
1.4	Used definitions, terms and abbreviations	5
1.5	References	5
1.6	Important note to the reader	6
1.7	General Acronyms and Abbreviations.	7
2	Implementation model	8
2.1	Application Data Flow Diagram	8
2.2	Functional definition of Application Entities	8
2.3	Sequencing of Real World Activities	9
3	AE Specifications	10
3.1	AE CT-Tomoscan DICOM Export Specification	10
3.1.1	Association Establishment Policies	10
3.1.2	Association Initiation Policy	10
3.1.3	Association Acceptance Policy	23
4	Communication Profiles	24
4.1	Supported Communication Stacks	24
4.2	TCP/IP Stack	24
4.2.1	Physical Media Support	24
5	Extensions/Specializations/Privatizations	24
6	Configuration	24
6.1	AE Title/Presentation Address mapping	24
6.1.1	Local AE Titles and Presentation Addresses	24
6.1.2	Remote AE Titles and Presentation Addresses	24
6.2	Configurable parameters	24
7	Support of Extended Character Sets	24

1 Introduction

This chapter provides general information about the purpose, scope and contents of this Conformance Statement.

1.1 Scope and field of application

The scope of this DICOM Conformance Statement is to facilitate data exchange with equipment of Philips Medical Systems. This document specifies the compliance to the DICOM standard (formally called the NEMA PS 3.X-1996 standards). It contains a short description of the applications involved and provides technical information about the data exchange capabilities of the equipment. The main elements describing these capabilities are: the supported DICOM Service Object Pair (SOP) Classes, Roles, Information Object Definitions (IOD) and Transfer Syntaxes.

The field of application is the integration of the Philips Medical Systems equipment into an environment of medical devices.

This Conformance Statement should be read in conjunction with the DICOM standard and its addenda [DICOM]. The conformance to the DICOM standard is a key element of the Inturis Program (see [INTURIS]).

1.2 Intended audience

This Conformance Statement is intended for:

- (potential) customers,
- system integrators of medical equipment,
- marketing staff interested in system functionality,
- software designers implementing DICOM interfaces.

It is assumed that the reader is familiar with the DICOM standard.

1.3 Contents and structure

The DICOM Conformance Statement is contained in chapter 2 through 7 and follows the contents and structuring requirements of DICOM PS 3.2-1996.

1.4 Used definitions, terms and abbreviations

DICOM definitions, terms and abbreviations are used throughout this Conformance Statement. For a description of these, see NEMA PS 3.3-1996 and PS 3.4-1996.

The word Philips in this document refers to Philips Medical Systems.

1.5 References

- [DICOM] The Digital Imaging and Communications in Medicine (DICOM) standard:
NEMA PS 3.X 1996
National Electrical Manufacturers Association (NEMA) Publication Sales
1300 N. 17th Street, Suite 1847
Rosslyn, Va. 22209, United States of America

[INTURIS] Inturis for Cardiology
On-Line Image Access
Doc. nr. 4522 982 69681
Philips medical Systems Ned. BV

1.6 Important note to the reader

This Conformance Statement by itself does not guarantee successful interoperability of Philips equipment with non-Philips equipment. The user (or user's agent) should be aware of the following issues:

- **Interoperability**

Interoperability refers to the ability of application functions, distributed over two or more systems, to work successfully together. The integration of medical devices into a networked environment may require application functions that are not specified within the scope of DICOM. Consequently, using only the information provided by this Conformance Statement does not guarantee interoperability of Philips equipment with non-Philips equipment. It is the user's responsibility to analyse thoroughly the application requirements and to specify a solution that integrates Philips equipment with non-Philips equipment.

- **Validation**

Philips equipment has been carefully tested to assure that the actual implementation of the DICOM interface corresponds with this Conformance Statement.

Where Philips equipment is linked to non-Philips equipment, the first step is to compare the relevant Conformance Statements. If the Conformance Statements indicate that successful information exchange should be possible, additional validation tests will be necessary to ensure the functionality, performance, accuracy and stability of image and image related data. It is the responsibility of the user (or user's agent) to specify the appropriate test suite and to carry out the additional validation tests.

- **New versions of the DICOM Standard**

The DICOM Standard will evolve in future to meet the user's growing requirements and to incorporate new features and technologies. Philips is actively involved in this evolution and plans to adapt its equipment to future versions of the DICOM Standard. In order to do so, Philips reserves the right to make changes to its products or to discontinue its delivery.

The user should ensure that any non-Philips provider linking to Philips equipment, also adapts to future versions of the DICOM Standard. If not, the incorporation of DICOM enhancements into Philips equipment may lead to loss of connectivity (in case of networking) and incompatibility (in case of media).

1.7 General Acronyms and Abbreviations.

The following acronyms and abbreviations are used in the document.

• ACC	American College of Cardiology
• AE	Application Entity
• ACR	American College of Radiology
• ANSI	American National Standard Institute
• BOT	Basic Offset Table
• CD-R	CD Recordable
• CD-M	CD Medical
• DCI	Digital Cardio Imaging
• DCR	Dynamic Cardio Review
• DICOM	Digital Imaging and Communication in Medicine
• DIMSE	DICOM Message Service Element
• DIMSE-C	DICOM Message Service Element-Composite
• DIMSE-N	DICOM Message Service Element-Normalized
• ELE	Explicit VR Little Endian
• EBE	Explicit VR Big Endian
• FSC	File Set Creator
• GUI	Graphic User Interface
• HIS	Hospital Information System
• HL7	Health Level Seven
• ILE	Implicit VR Little Endian
• IOD	Information Object Definition
• ISIS	Information System - Imaging System
• NEMA	National Electrical Manufacturers Association
• PDU	Protocol Data Unit
• RIS	Radiology Information System
• RWA	Real World Activity
• SC	Secondary Capture
• SCM	Study Component Management
• SCP	Service Class Provider
• SCU	Service Class User
• SOP	Service Object Pair
• TCP/IP	Transmission Control Protocol/Internet protocol
• UID	Unique Identifier
• WLM	Worklist Management

2 Implementation model

The CT Tomoscan M/EG R 3.1 (short CT-Tomoscan) system of Philips Medical Systems is a Computed Tomography (CT) image generating system. The system can be installed with an Export function based on the DICOM Image Storage to transfer images and image related data from the CT-Tomoscan to a remote system. This DICOM Export function implies the presence of the CT Tomoscan software release 2.6 or higher.

The DICOM Export function is described in this document.

2.1 Application Data Flow Diagram

The CT-Tomoscan system behaves as a single Application Entity. The related Implementation Model is shown in Figure 2-1 on page 8.

The DICOM Export function can be activated by the CT-Tomoscan operator. The images to be sent are selected from the images of several acquisition runs of the same patient (also called examination). Each image in these runs can be flagged for export or not. At export request the flagged images will be converted into DICOM format and sent out to a remote destination. This destination is previously selected by the operator from the user interface.

Image data transferred are DICOM Computed Tomography (CT) Class instances.

The images transferred are intended for viewing purposes. Postprocessing like MPR, 3D reconstruction and rendering may be possible, depending on the capabilities of the workstation receiving the CT images.




Figure 2-1: CT-Tomoscan DICOM Export and Print Implementation

2.2 Functional definition of Application Entities

The CT-Tomoscan AE acts as a Service Class User (SCU) of the Storage Service Class. The export function acts on the selected images of one examination. When the export is initiated, the AE will open an association to the remote system. The selected images and related image data are converted into a DICOM message to be sent to the remote system.

The application does not support Storage Service Classes as Service Class Provider (SCP).

2.3 Sequencing of Real World Activities

Not applicable.

3 AE Specifications

3.1 AE CT-Tomoscan DICOM Export Specification

The CT-Tomoscan DICOM Export Application Entity provides Standard Conformance to the following DICOM V3.0 SOP Classes as an SCU:

Table 3-1: Supported SOP Classes by the CT-Tomoscan DICOM Export AE as SCU

SOP Class Name	UID
CT Image Storage	1.2.840.10008.5.1.4.1.1.2
Basic Grayscale Print Management Meta SOP Class	1.2.840.10008.5.1.1.9
> ^a Basic Film Session SOP Class	1.2.840.10008.5.1.1.1
> Basic Film Box SOP Class	1.2.840.10008.5.1.1.2
> Basic Grayscale Image Box SOP Class	1.2.840.10008.5.1.1.4
> Printer SOP Class	1.2.840.10008.5.1.1.16

a. The '>' sign indicates that the SOP Class is part of the above mentioned Meta SOP Class.

The CT-Tomoscan Application Entity does not support DICOM V3.0 SOP Classes as an SCP.

3.1.1 Association Establishment Policies

3.1.1.1 General

CT-Tomoscan has a configurable maximum PDU size in steps (default is 16K = 16384 bytes).

3.1.1.2 Number of Associations

CT-Tomoscan will attempt to establish one association at a time.

3.1.1.3 Asynchronous Nature

CT-Tomoscan does not support asynchronous operations and will not perform asynchronous window negotiation.

3.1.1.4 Implementation Identifying Information

The Implementation Class UID is: "1.3.46.670589.10.13.800143.1"

The implementation version name is: "HMC/M_TOMO_20"

3.1.2 Association Initiation Policy

CT-Tomoscan initiates associations as a result of the following local Real-World activities:

- The Image Export Request to send the selected images from the CT-Tomoscan to a remote system.

3.1.2.1 Request to send images from CT-Tomoscan to a remote system**3.1.2.1.1 Associated Real-World Activity**

After selection of a peer station and after selection of one of more images of a number of scans, these images will be sent when initiating the Send command. CT-Tomoscan initiates one association to the selected peer entity and uses it to send the selected images via C-STORE requests (and receives the associated C-STORE Responses). The association is released by CT-Tomoscan after successful transfer of the images or when an error occurs. CT-Tomoscan handles each send request one after another.

3.1.2.1.2 Proposed Presentation Contexts

CT-Tomoscan will propose the following presentation contexts:

Table 3-2: Proposed Presentation Contexts for Send Request

Presentation Context table					
Abstract Syntax		Transfer Syntax		Role	Extended Negotiation
Name	UID	Name List	UID List		
CT Image Storage - STORE	1.2.840.10008.5.1.4.1.1.2	ILE	1.2.840.10008.1.2	SCU	None
CT Image Storage - STORE	1.2.840.10008.5.1.4.1.1.2	ELE	1.2.840.10008.1.2.1	SCU	None

3.1.2.1.3 C-STORE SCU Conformance

CT-Tomoscan has the following behaviour on successful (with or without warnings) and unsuccessful transfer of images:

- Success (return status 0000)
The successful transfer is indicated on the console: 'Done'.
- Refused (return status A7xx), Error (return status A9xx or Cxxx) and Warning (return status B00x)
The failed transfer is indicated on the console: 'Network Error'. The reason is not shown.

While busy with transfer, the status Busy is shown on the CT-Tomoscan console.

Extended negotiation is not supported.

Table 3-3 lists the applied Conditional (DICOM Type 1C and 2C) and Optional (DICOM Type 3) attributes of the standard CT Image IOD. These attributes may be empty. Chapter 8 on page 19 gives an overview of all modules and attribute details.

Table 3-3: Applied Conditional and Optional Attributes of the CT Image IOD

Information Entity	Module	Conditional Attributes	Optional Attributes
Patient	Patient	-	-
Study	General Study	-	-
	Patient Study	-	Admitting Diagnosis Description, Patient's Age
Series	General Series	Patient Position	Series Date, Series Time, Performing Physician's Name, Operators' Name, Protocol Name
Frame of Reference	Frame of Reference	-	-
Equipment	General Equipment	-	Institution Name, Manufacturer's Model Name
Image	General Image	Image Date, Image Time	Referenced Image Sequence, Image Comments
	Image Plane	-	Slice Location
	Image Pixel	-	-
	Contrast/Bolus	-	-
	CT Image	-	Scan Options, Reconstruction Diameter, Gantry/Detector Tilt, Exposure Time, X-Ray Tube Current, Convolution Kernel
	VOI LUT	-	Window Center, Window Width
	SOP Common	Specific Character Set	-

3.1.2.1.4 Overview applied CT Image IOD

The modules selected from the CT Image IOD module table of DICOM 3.0 are given in the table below.

Table 3-4: Applied Modules in the CT Image IOD

Information Entity	Module
Patient	Patient
Study	General Study
	Patient Study
Series	General Series

Table 3-4: Applied Modules in the CT Image IOD (Continued)

Information Entity	Module
Frame of Reference	Frame of Reference
Equipment	General Equipment
Image	General Image
	Image Plane
	Image Pixel
	Contrast/Bolus
	CT Image
	VOI LUT
	SOP Common

The details of these applied modules are given in the tables below. The list of possible values are given (if applicable). The situation that an attribute is present conditionally/optionally or that an attribute may contain a zero length value, is indicated too. Conditions and Defined/Enumerated Values of DICOM 3.0 are applicable but are not always shown.

Table 3-5: CT Image Storage SOP Class - Patient Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Patient's Name	0010,0010	
Patient ID	0010,0020	
Patient's Birth Date	0010,0030	
Patient's Sex	0010,0040	Applied value(s): F, M, O

Table 3-6: CT Image Storage SOP Class - General Study Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Study Date	0008,0020	
Study Time	0008,0030	
Accession Number	0008,0050	
Referring Physician's Name	0008,0090	
Study Instance UID	0020,000D	

Table 3-6: CT Image Storage SOP Class - General Study Module (Continued)

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Study ID	0020,0010	

Table 3-7: CT Image Storage SOP Class - Patient Study Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Admitting Diagnosis Description	0008,1080	May be empty
Patient's Age	0010,1010	

Table 3-8: CT Image Storage SOP Class - General Series Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Series Date	0008,0021	
Series Time	0008,0031	
Modality	0008,0060	Applied value(s): CT
Performing Physician's Name	0008,1050	Empty if not filled in
Operators' Name	0008,1070	Empty if not filled in
Protocol Name	0018,1030	
Patient Position	0018,5100	Applied value(s): FFDL, FFDR, FFP, FFS, HFDL, HFDR, HFP, HFS
Series Instance UID	0020,000E	
Series Number	0020,0011	Value 0 for Scannograms, value 1,2,3... for Slices

Table 3-9: CT Image Storage SOP Class - Frame of Reference Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Frame of Reference UID	0020,0052	
Position Reference Indicator	0020,1040	

Table 3-10: CT Image Storage SOP Class - General Equipment Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Manufacturer	0008,0070	Applied value(s): Philips Medical Systems
Institution Name	0008,0080	Empty if not filled in
Manufacturer's Model Name	0008,1090	Applied value(s): CT Tomoscan M-EG

Table 3-11: CT Image Storage SOP Class - General Image Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Image Date	0008,0023	
Image Time	0008,0033	
Referenced Image Sequence	0008,1140	Present for Axial images, pointing to the related Scannogram
> Reference SOP Class UID	0008,1150	
> Reference SOP Instance UID	0008,1155	
Image Number	0020,0013	
Image Comments	0020,4000	Empty if not filled in

Table 3-12: CT Image Storage SOP Class - Image Plane Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Slice Thickness	0018,0050	
Image Position (Patient)	0020,0032	
Image Orientation (Patient)	0020,0037	
Slice Location	0020,1041	
Pixel Spacing	0028,0030	

Table 3-13: CT Image Storage SOP Class - Image Pixel Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Samples per Pixel	0028,0002	See CT Image Module
Photometric Interpretation	0028,0004	See CT Image Module
Rows	0028,0010	
Columns	0028,0011	
Bits Allocated	0028,0100	See CT Image Module
Bits Stored	0028,0101	See CT Image Module
High Bit	0028,0102	See CT Image Module
Pixel Representation	0028,0103	
Pixel Data	7FE0,0010	

Table 3-14: CT Image Storage SOP Class - Contrast/Bolus Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Contrast/Bolus Agent	0018,0010	May be empty

Table 3-15: CT Image Storage SOP Class - CT Image Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Image Type	0008,0008	Applied value(s): ORIGINAL, DERIVED \ PRIMARY, SECONDARY \ AXIAL, LOCALIZER
Scan Options	0018,0022	
KVP	0018,0060	
Reconstruction Diameter	0018,1100	
Gantry/Detector Tilt	0018,1120	
Exposure Time	0018,1150	
X-Ray Tube Current	0018,1151	
Convolution Kernel	0018,1210	
Acquisition Number	0020,0012	

Table 3-15: CT Image Storage SOP Class - CT Image Module (Continued)

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Samples per Pixel	0028,0002	
Photometric Interpretation	0028,0004	Applied value(s): MONOCHROME1, MONOCHROME2
Bits Allocated	0028,0100	Applied value(s): 16
Bits Stored	0028,0101	Applied value(s): 12
High Bit	0028,0102	Applied value(s): 11
Rescale Intercept	0028,1052	
Rescale Slope	0028,1053	

Table 3-16: CT Image Storage SOP Class - VOI LUT Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Window Center	0028,1050	
Window Width	0028,1051	

Table 3-17: CT Image Storage SOP Class - SOP Common Module

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Specific Character Set	0008,0005	Applied value(s): ISO_IR 100
SOP Class UID	0008,0016	Applied value(s): 1.2.840.10008.5.1.4.1.1.2
SOP Instance UID	0008,0018	

3.1.2.2 Print Images

3.1.2.2.1 Associated Real-World Activity

There are two ways to request for image printing:

- Print Compose

The operator is able to select one or more images from the internal database (via the Data Handling facility) and perform the Print operation on them.

- Print Protocol

The operator is also able to print images via the various clinical applications of CT-Tomoscan.

The operator will select the print destination (out of choice list of configured printers) and some print parameters (depending on the configuration and the selected printer).
As a result, CT-Tomoscan will initiate an association to the selected printer and uses it to send the Print Service Elements of the Print SOP Classes.

CT-Tomoscan allows to have a print preview first.

3.1.2.2.2 Proposed Presentation Context

CT-Tomoscan will propose the presentation contexts as given in: Table 3-3, “Proposed Presentation Contexts for EasyVision to Other,” on page 15, for the Basic Grayscale Print Management Meta SOP Class (UID 1.2.840.100008.5.1.1.9).

3.1.2.2.3 Conformance to the Print SOP Classes

CT-Tomoscan provides standard conformance to the Basic Grayscale Print Management Meta SOP Class.

The applied order of Print Service Elements (DIMSE's) is specified in Table 3-18. A description and the applied optional (i.e. non-mandatory attributes as Print SCU) attributes in these Service Elements are specified too. Note that the Service Elements order is not specified by the DICOM standard.

An explicit N-DELETE Request on the created instances is not done by CT-Tomoscan; these are deleted implicitly when releasing the association.

The **full list of (Mandatory and Optional) attributes** applied in these Service Elements are given in section 8 on page 36.

Table 3-18: The applied order of Print Service Elements and its optional attributes

Service Element of SOP Class	Description and applied optional attributes
N-GET of the Printer SOP Class	Purpose is to retrieve printer information.
N-CREATE of the Basic Film Session SOP Class	CT-Tomoscan specifies the DICOM Printer about some general presentation parameters, applicable for all films in the Film Session. Applied optional attributes are: Number of Copies, Print Priority, Medium Type, Film Destination

Table 3-18: The applied order of Print Service Elements and its optional attributes

Service Element of SOP Class	Description and applied optional attributes
N-CREATE of the Basic Film Box SOP Class	CT-Tomoscan specifies the DICOM Printer about some general presentation parameters, applicable for all images in the Film Box. Applied optional attributes are: Film Orientation, Film Size ID, Magnification Type, Max. Density, Configuration Information, Trim.
N-SET of the Basic Gray-scale Image Box SOP Class	CT-Tomoscan will send the images to be printed. Applied optional attributes are: Polarity
N-ACTION of the Basic Film Box SOP Class	CT-Tomoscan triggers the DICOM Printer to print, this actual print action is done at film box level. No (optional) attributes are present.

The table below specifies the supported Service Elements which may be generated by the Printer at any time during the association.

Table 3-19: The applied sequence of Print Service Elements and its optional attributes

Service Element of SOP Class	Note
N-EVENT-REPORT of the Printer SOP Class	May be sent at any moment by the Printer SCP (i.e. the DICOM Printer). CT-Tomoscan will respond.

The Status Codes of DIMSE Responses (Success, Warning, Failure) as returned by the printer will also be logged (for service purposes) and are mapped onto??? **Provide information!**

The following implementation remarks are important to achieve successful printing:

- CT-Tomoscan will release the association when the print command is given (i.e. the N-ACTION Request); the association is not kept open for receiving N-EVENT-REPORTs of the Printer SOP Class.

3.1.2.2.4 Basic Film Session SOP Class

Table 3-20: Basic Film Session SOP Class - N-CREATE

Attribute Name	Tag	Note
Number of Copies	2000,0010	VR = IS (Integer String; max 12), VM = 1 Between 1 and 99.

Table 3-20: Basic Film Session SOP Class - N-CREATE (Continued)

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Print Priority	2000,0020	VR = CS (Code String; max 16), VM = 1 Applied value(s): High
Medium Type	2000,0030	VR = CS (Code String; max 16), VM = 1 Applied value(s): BLUE FILM, CLEAR FILM, PAPER
Film Destination	2000,0040	Applied Values(s): PROCESSOR, MAGAZINE

3.1.2.2.5 Description of the Basic Film Box SOP Class**Table 3-21: Basic Film Box SOP Class - N-CREATE**

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Image Display Format *	2010,0010	VR = ST (Short Text; max 1024), VM = 1 The applied value below is an CT-Tomoscan specific value indicating that one (large) image is contained in a Film Box. Applied value(s): STANDARD\1,1 (Rows, Columns)
Film Orientation *	2010,0040	VR = CS (Code String; max 16), VM = 1 Applied value(s): LANDSCAPE, PORTRAIT
Film Size ID *	2010,0050	VR = CS (Code String; max 16), VM = 1 DICOM specifies a number of Defined Terms; more values are possible and is print configuration dependent.
Magnification Type	2010,0060	VR = CS (Code String; max 16), VM = 1 Normally sent out, however some DICOM printers are not able to handle (value NONE for) this attribute. Applied value(s): CUBIC
Trim	2010,0140	VR = CS (Code String; max 16), VM = 1 Applied value(s): NO
Configuration Information	2010,0150	VR = ST (Short Text; max 1024), VM = 1 Used to refer to a vendor specific Lookup-table (LUT); should be applied by the DICOM printer if LUT data is present.
Referenced Film Session Sequence	2010,0500	VR = SQ (Sequence of Items; -), VM = 1 Parent Film Session.
> Referenced SOP Class UID	0008,1150	VR = UI (Unique Identifier; max 64), VM = 1
> Referenced SOP Instance UID	0008,1155	VR = UI (Unique Identifier; max 64), VM = 1

* Note: that Image Display, Film Orientation and Film Size ID are selectable. Which combination can be selected is dependable on the specific Printer Type.

Table 3-22: Basic Film Box SOP Class - N-ACTION

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
No attributes present		

3.1.2.2.6 Basic Grayscale Image Box SOP Class

Table 3-23: Basic Grayscale Image Box SOP Class - N-SET

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Image Position	2020,0010	VR = US (Unsigned Short; 2), VM = 1 Applied value(s): 1
Polarity	2020,0020	VR = CS (Code String; max 16), VM = 1 Applied value(s): NORMAL
Preformatted Grayscale Image Sequence	2020,0110	VR = SQ (Sequence of Items; -), VM = 1
> Samples per Pixel	0028,0002	VR = US (Unsigned Short; 2), VM = 1 Applied value(s): 1
> Photometric Interpretation	0028,0004	VR = CS (Code String; max 16), VM = 1 Applied value(s): MONOCHROME2
> Rows	0028,0010	VR = US (Unsigned Short; 2), VM = 1 Depending on the selected printer type and film size
> Columns	0028,0011	VR = US (Unsigned Short; 2), VM = 1 Depending on the selected printer type and film size
> Bits Allocated	0028,0100	VR = US (Unsigned Short; 2), VM = 1 Applied value(s): 16, 8
> Bits Stored	0028,0101	VR = US (Unsigned Short; 2), VM = 1 Applied value(s): 12, 8
> High Bit	0028,0102	VR = US (Unsigned Short; 2), VM = 1 Applied value(s): 11, 7
> Pixel Representation	0028,0103	VR = US (Unsigned Short; 2), VM = 1 Applied value(s): 0x0000 (i.e. unsigned integer)
> Pixel Data	7FE0,0010	VR = OW (Other Word String; -) or OB (Other Byte String; -), VM = 1

3.1.2.2.7 Printer SOP Class

Check is this paragraph is correct.

Table 3-24: Printer SOP Class - N-GET


<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Printer Status	2110,0010	VR = CS (Code String; max 16), VM = 1
Printer Status Info	2110,0020	VR = CS (Code String; max 16), VM = 1

Table 3-25: Printer SOP Class - N-EVENT-REPORT^a

<i>Attribute Name</i>	<i>Tag</i>	<i>Note</i>
Printer Status Info	2110,0020	VR = CS (Code String; max 16), VM = 1 Conditionally sent by the Printer. CT-Tomoscan will react on notification, and will show corresponding messages on the UI for the events WARNING and FAILURE.

a. This Service Element is sent by the printer and interpreted by CT-Tomoscan.

The CT-Tomoscan does not send an attribute list to the printer, therefore the only attributes which are needed to be supported by the printer, are the mandatory attributes listed in Table 3-24, "Printer SOP Class - N-GET," on page 22.

**Figure 3-1: CT-Tomoscan general behaviour for N_GET and N_EVENT_REPORT**

3.1.3 Association Acceptance Policy

The CT-Tomoscan Application Entity does not accept associations.

4 Communication Profiles

4.1 Supported Communication Stacks

CT-Tomoscan provides DICOM V3.0 TCP/IP Network Communication Support as defined in Part 8 of the DICOM Standard.

4.2 TCP/IP Stack

CT-Tomoscan uses the TCP/IP program installed on the PC where CT-Tomoscan DICOM Export is running on. This is a subroutine library that is based on a Berkeley socket interface.

4.2.1 Physical Media Support

CT-Tomoscan supports ISO 8802-3 10BASE5 (Thick-wire), 10BASE2 (Thin-wire) and 10BASE-T (Twisted-pair) Ethernet.

5 Extensions/Specializations/Privatizations

None.

6 Configuration

The CT-Tomoscan system is configured by means of editing the DICOM configuration files on the system running the DICOM Export. This configuration is intended to be used by Philips service engineers only.

6.1 AE Title/Presentation Address mapping

6.1.1 Local AE Titles and Presentation Addresses

The CT-Tomoscan AE title is not configurable; the system chooses an AE title which is unique within a network. This AE title can be seen via the System Administration function.

6.1.2 Remote AE Titles and Presentation Addresses

For remote applications that act as Service Class Provider the following additional information must be provided:

- The AE title.
- The host name and IP address on which the application resides.
- The port number at which the application accepts association requests.

6.2 Configurable parameters

The maximum PDU size is configurable in steps of 1K, 2K, 4K, 8K and 16K.

7 Support of Extended Character Sets

CT-Tomoscan supports the Extended Character Set "ISO_IR 100" which is the Latin alphabet No 1, supplementary set.

